

Konformitätserklärung (ISO/IEC 17050-1)
Testkörper
Supplier's declaration of conformity (in accordance with ISO/IEC 17050-1)
Test sample

Werkstoff Material	Durchmesser Diameter	Ident-Nr. Ident-no.	Dichte Density	Maß-/Formgenauigkeit Dimensional accuracy DIN 5401-2002-08	Magnetische Permeabilität Magnetic permeability
Ferrous 1.3505, AISI 52100	0,80 mm	IK10063	7,8g/cm ³	Grade 5	--
	1,00 mm	IK10051		Grade 28	--
		IK10080			--
	1,20 mm	IK10055		Grade 28	--
	1,50 mm	IK10065		Grade 28	--
		IK10081			--
	1,80 mm	IK10047		Grade 28	--
	2,00 mm	IK10045		Grade 28	--
		IK10056			--
	2,50 mm	IK10059		Grade 28	--
	3,00 mm	IK10062		Grade 28	--
	4,00mm	IK10077		Grade 28	--
	4,50 mm	IK10029		Grade 28	--
	5,00 mm	IK10030		Grade 28	--
	5,50 mm	IK10031		Grade 28	--
	6,00mm	IK10032		Grade 28	--

Lebensmittelbescheinigung Testkörperummantelung
Food certificate test ball cover
☒ FDA-Konformitätserklärung / FDA-Declaration of conformity

Hiermit bestätigen wir, dass die Ummantelung unserer Testkörper aus Kunststoffen/Elastomeren, die bei bestimmungsgemäßem Gebrauch in Kontakt mit Lebensmitteln kommen können, die allgemeinen Anforderungen der FDA-Richtlinie 21 CFR §177 erfüllen.

We hereby certify that our test sample covers made of plastics/elastomers, which could come into contact with foodstuffs in the course of their intended use, meet the requirements under FDA directive 21 CFR Section 177

Enthaltene Additive entsprechen jeweils den zutreffenden FDA-Anforderungen.

Any additives that they contain meet the respectively applicable FDA requirements:

POM Blau / POM Blue (TestSticks, Testballs) - CFR 21, Part 177.2470 (Polyoxymethylene copolymer) - CFR 21, Part 178.3297 (colorants)	PE / PE (FlexStick Schlauch/FlexStick hose) - CFR 21, Part 177.1520 (Olefin polymers)	EVA Co-polymer (Flyball) - CFR 21, Part 177.1350 (EVA Co-polymers)
CARD LINE XD (Xray-Cards/Teststripes) - CFR 21, part 177.1580 (Polycarbonate)		

Sesotec GmbH

02.11.2023



 i. V. Dieter Sprenzinger
 - Manager Quality -